

Centre for Cellular and Molecular Platforms

(A Dept. of Biotechnology, Govt. of India Initiative)

GKVK, Bellary Road, Bangalore 560 065. India

Tel. No: 080 2366 6346

Email: nalinir@ncbs.res.in, www.ncbs.res.in

ENQUIRY

Ref: C-Camp-L-058/2019-20(I)

04 December 2019
=>By Speed Post/Courier

Dear Sirs,

Please let us have your Quotation for the following:

Sl.No	Item Description	Qty
1	<p><u>Specifications for: BOD Incubator</u></p> <p><u>Chamber Design:</u> Double walled</p> <p><u>Inner Chamber:</u> Inner chamber should be made of stainless steel 304 Grade</p> <p><u>Outer Chamber:</u> Outer chamber with mild steel and powder coated exteriors.</p> <p><u>Insulation:</u> Insulation between inner and outer chamber filled with high grade glass wool to withstand maximum Temperature 60° C</p> <p><u>Wheels:</u> The unit should be mounted on a sturdy steel frame with cast iron wheels for easy movement</p> <p><u>Conditioning system Air Circulation Fan Motor:</u> The triple walled back of the instrument should be provided with air circulation fans. Temperature and air circulation uniformity should be maintained.</p> <p><u>Cooling:</u> Energy efficient cooling unit with hermetically sealed compressors. Noise free, vibration free and air cooled Compressors. Start in sequence to protect the unit from overloading. CFC free environmentally friendly refrigerant and compressors of Kirloskar/Tecumseh make, conforming to latest international standards and guidelines</p> <p><u>Inner Glass Door:</u> System should have inner full- length glass door to view the samples without disturbing the temperature</p> <p><u>Outer Door:</u> Inner stainless steel and Outer made of Mild steel finished with powder coated</p> <p><u>Gasket:</u> Rubber gasket and latch is provided for the door is provided for the door to avoid air leakage and temperature loss from Instrument.</p> <p><u>Trays:</u> Three perforated removable stainless steel trays.</p> <p><u>Temperature Range:</u> 10°C to 60°C</p> <p><u>Temperature Controller:</u> Microprocessor based PID temperature controller with dual display Process value and set value), low and high limit setting with PT100 sensor and Timer control (optional).</p> <p><u>Temperature Resolution:</u> 0.1°C</p> <p><u>Temperature Accuracy:</u> ± 2°C</p> <p><u>Temperature Uniformity:</u> ±3°C.</p> <p><u>Power Supply:</u> 230 Volts Single Phase 50Hz, AC supply.</p> <p><u>Inner viewing Light:</u> Door switch with viewing to be provided without additional cost</p> <p><u>Chamber Illumination:</u> Cyclic Timer with florescent tube / LED light to be provided on either side of the inner Chamber without additional cost.</p>	01. No.

NOTE: 1. Please attach Authorization letter from your Principle supplier with quotation.
2. IMPORTANT NOTE: C-Camp is Initiative of Dept. of Biotechnology, Ministry of Science and Technology, Govt of India and is entitled to concessional rate of GST @ 5% for items supplied for research purpose vide notification no. 45/2017 dated 14th Nov, 2017. The offer should be submitted after fully considering the above notification.
3. The Price quoted against this RFQ should be extended to Bangalore Life Science Cluster (BLiSc) of NCBS, Instem and C-CAMP for placing repeat order as per norms, by any one or all institutes of Bangalore Life Science Cluster (BLiSc)

- 1. The bids shall be enclosed in an envelope and due date sealed duly marked Tender for Ref No: C-Camp-L-058/2019-20 (I) The bids should be addressed and to be mailed to 'THE HEAD-PURCHASE'.** The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and C-Camp will not be responsible for the delay.
- 2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 24/12/2019**
- 3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.**
4. The validity of your quotation should be for 60 days from the due date.
5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.
6. Your quotation should indicate delivery period & warranty period.
7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done up to C-Camp Stores.
8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.
9. Payment: within one month after delivery & acceptance/satisfactory installation.
10. Please ensure that the enquiry number and the due date is super scribed on the envelope failing which your quotations liable to be rejected.
11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except advalorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No.51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipment's, technical instruments, equipment's (including computers), their accessories, spares, consumables and software. Hence, please offer your prices
- 12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with Xerox copy of the DGS&D order.**
13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.
- 14. The quotations may be submitted in USD, Euro, GBP, JPY and any tradable currency.**
15. Liquidity Damages: If the equipment/ items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.
16. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services under section 194J, the tax reduction shall be at the rate of 10%. In case service provider does not provide PAN number, the deduction shall be at 20% under section 206 AA. -Tax Deduction Certificates will be provided on request for non-PAN holders & Foreign Vendors and PAN holders could avail them through NSDL site dealing with 26AS.

Yours faithfully
 For and on behalf of
 Centre for Cellular and Molecular Platforms

NALINI R