

Ref: C-CAMP/L-127/2023-24 (C)

January 03, 2024

ENQUIRY BUY BACK


Dear Sirs,

Please let us have your lowest quotation for the following:

Sl.No.	Item/s description	Qty.
1	<p>Ultra Low Temperature, Vertical Upright type -86° C Deep Freezer</p> <p>Operating temperature range should be from -70°C to -86°C at ambient temperature of 35-degree C. System should be vertical standing and should have working volume of 550 Ltr (± 5%) System should have minimum four shelves with individual inner SS 304 doors with gasket and independent door hinges and latches. System should be equipped with precise temperature controller with accuracy of ± 2°C or better & it should maintain uniform temperature in freezer. System should have pressure equalization port (PEP). System should be fully programmable microprocessor controlled with keypad and suitable touch control panel. System should have digital display with resistive touch control for all functions (even touching with the gloves). The digital display should be clear, easy to read LCD/LED display and USB for exporting the data. System should have onboard diagnostics digital display that allows for quicker diagnosis of problem minimizing downtime. Inbuilt on-board diagnostic digital Log data storage for Temperature and alarms. Data must also be downloadable via a USB port. System should have low/high voltage protection, surge protection to minimize potential damage from voltage spikes. System should be equipped with buzzer and visual alarm for power failure and temperature deviation against the set temperature or any temperature deviation should trigger audible and visual warning. The Freezer shall have minimum two- 1-inch access ports. The Freezer door must open at least 180 degrees for easy sample access. The Freezer shall allow set-point security control that blocks specific users from changing freezer set point or alarms through the use of a user name and password control and should allow multiple users. The system should record the temperature and history of the freezer for life time. System should have robust and silent compressor with noise level less than 45 db. The freezer door recovery time should be less than 10 minutes (± 10%) System should have advanced cooling system with high power efficient compressor units like inverter type or dual cooling system. System should have HFC Free Natural refrigerants compliant with EU F-Gas regulation and certification must be enclosed. System should have connectivity port for BMS control for all the parameters required for monitoring including NO-NC contacts. System should have enhanced acoustic insulation. The system should meet the international safety standards and CE, ISO 9001, ISO13485, USFDA Marked or equivalent marked. System should be offered with 5-year OEM warranty as part of offer (not as extra item). System should have door handle and latch assembly with standard key lock. Offer should have reference of minimum 5 similar research institute for installation and satisfactory performance and should be submitted with the quotation. System should be suitable for power rating 230v, 50Hz single phase. System should have battery backup for controller, display and alarm. System should be highly energy efficient. The average power consumption of the system per day at normal working condition to be given. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required). Certificate of calibration and inspection from factory to be supplied with system. System should be provided with suitable 5KVA servo voltage stabilizer. The system should be installed by the trained Engineer and users training should be given on application, operation and maintenance. The service team should be in Bangalore. Contact details of Engineer should be submitted with this offer. The past performance/service support in NCBS/inStem/CCAMP should be satisfactory and the technical evaluation will be done accordingly. The system should be supplied with all accessories required for function of the equipment. The offer should include a buyback proposal for old -80deg freezer at C-CAMP.</p>	02
	BUY BACK ITEM DETAILS	
1	MDF-U54V -86 Deg C Deep Freezer, 230V/50Hz, Capacity 519L, FOOT	1 No.
2	MDF-U74V -86Deg C Deep Freezer, 230V/50Hz, Capacity 728L	2 Nos.

Note:

1. The quotation shall be submitted in a sealed envelope duly superscribed with the enquiry number, and the due date for Ref No:..... The bids should be addressed and to be sent by post/courier to 'THE HEAD-PURCHASE'.
- The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and CCAMP/NCBS will not be responsible for the delay.
- 2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 17/01/2024 BY 5.30PM.**
3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.
4. The validity of your quotation should be for 60 days from the due date.
5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.
6. Your quotation should indicate delivery period & warranty period.
7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto CCAMP Stores.
8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.
9. Payment: within one month after delivery & acceptance/satisfactory installation.
10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.
- 11. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.**
12. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.
13. Since we are a research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No.51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices.
14. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.
15. CCAMP is a public funded research institute and is entitled to concessional rate of GST @ 5% for certain items supplied for research purpose vide notification no. 47/2017 and 45/2017 dated 14th Nov, 2017. The offer should be submitted after fully considering the above notification.
16. Liquidity Damages: If the equipment/ items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.
17. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services



Chetana S. R.

GKVK, Bellary Road, Bangalore 560 065, India
Phone +91-80-23666344. Fax +91-80-23636662
chetana@ncbs.res.in . www.ccamp.res.in