

Ref: C-CAMP/L-135/2021-22 (C)

September 22, 2021

ENQUIRY

Dear Sirs,

Please let us have your lowest quotation for the following:

Sl.No.	Item/s description	Qty.
1	<p>Specification for Stackable Incubator shaker Two Units of Stackable programmable shaker incubatorwith capacity to accommodate flasks up to 6 L Should have Temperature range from 5°C above ambient to 80°C PI Microprocessor control with platinum RTD Sensor. Speed should be:25- 400 rpm,± 2 RPM Should have Accuracy and control: ±0.1°C at 37°C. Shaker should be equipped with single knob to selects all operating conditions quickly with large LCD display. Large viewing window and internal light provides easyviewing of samples without the need to open the door. Capable of calibrating temperature and speed via keypad. At least 3 onboard programs should be there in shaker each program should have different time based steps to control temperature and speed.(Minimum 15 steps per program) Shaker should have drip pan. Should have built-in water reservoir to humidify chamber to reduce sample evaporation. Shaker with minimum foot print will be preferred.Dimensions (W × D × H) excluding base63.5 × 74.5 × 81.8 cm (25.0 × 29.3 × 32.2 in) or better. Humidity monitor, gassing manifold, remote alarm should be available optionally. Alarm must be activated if speed deviates± 5rpm. Shaker must have option of 19mm & 25mm drive orbit. Should come with Automatic Restart: Starts automatically after door opening and closing or power failure. Should have heavy duty counterbalanced Triple-eccentric counterbalanced drive in cast iron housing with 9 permanently lubricated ball bearings, Driven by solid-state brushless DC motor. Solid state, DC brushless motor Should have data logging via RS 232 and remote computer running conventional lab management software or hyper terminal. Should have acceleration circuit to prevents sudden start and stop to avoid cell damaging. Should have universal platform with dimension of 18"Wide X 18" Deep (46cm x 46cm). Should come with one stationary shelf for incubation as standard. Clamps starter set containing following clamps should be supplied with system 4 each 250 mL 4 each 500 mL 2 each 1000 mL 2 each 2000 mL Stacking kit to stack two incubators.WARRANTY : 5 YERAS</p>	02

Note:

1. The quotation shall be submitted in a sealed envelope duly superscribed with the enquiry number, and the due date for Ref No:..... The bids should be addressed and to be mailed / couriered (sent by post/courier) to "THE HEAD-PURCHASE".

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and CCAMP/NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 13/10/2021 BY 5.30PM.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto CCAMP Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

12. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties


13. Since we are a research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No.51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

14. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

15. CCAMP is a public funded research institute and is entitled to concessional rate of GST @ 5% for certain items supplied for research purpose vide notification no. 47/2017 and 45/2017 dated 14th Nov, 2017. The offer should be submitted after fully considering the above notification.

16. Liquidity Damages: If the equipment/ items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.

17. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services



Chetana S. R.

GKVK, Bellary Road, Bangalore 560 065. India
Phone +91-80-23666344. Fax +91-80-23636662
chetana@ncbs.res.in . www.ccamp.res.in