

Ref: C-CAMP/L-094/2021-22(C)

August 13, 2021

ENQUIRY

Dear Sirs,

Please let us have your lowest quotation for the following:

Sl.No.	Item/s description	Qty
1	Supply and installation of CO2 gas detector – as per below specifications	3 Nos.

Note:

1. The quotation shall be submitted in a sealed envelope duly superscribed with the enquiry number, and the due date for Ref No:..... The bids should be addressed and to be mailed / couriered (sent by post/courier) to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and CCAMP/NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 30/08/2021 BY 5.30PM.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto CCAMP Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

12. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties

13. Since we are a research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

14.If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

15. CCAMP is a public funded research institute and is entitled to concessional rate of GST @ 5% for certain items supplied for research purpose vide notification no. 47/2017 and 45/2017 dated 14th Nov, 2017. The offer should be submitted after fully considering the above notification.

16. Liquidity Damages: If the equipment/ items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.

17. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services

Chetana
13/8/21
Chetana S. R.

GKVK, Bellary Road, Bangalore 560 065. India
Phone +91-80-23666344. Fax +91-80-23636662
chetana@ncbs.res.in . www.ccamp.res.in

Specification for CO2 gas detector

- 1) System should have CO2 Measuring range 0-9999 PPM
- 2) System should have CO2 Measuring accuracy of ± 30 PPM
- 3) The system should have digital display and user-friendly control system
- 4) System should have Operating Temperature range 10°C - 40°C
- 5) System should generate audio alarm when there is deviation in set point, preferably there should be an option to mute the audible alarms
- 7) System should have large, easy to read LCD display.
- 8) System should have Li-ion battery for long term operation.
- 9) System should have an electrical adaptor to charge the battery, the operating voltage of adaptor is 230V, 50 Hz.
- 10) The system should have battery backup of 8 hours per day.
- 11) System should be supplied with all the accessories required to function.
- 12) Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).
- 13) The service team should be in Bangalore. Contact details of Engineer should be submitted with this offer
- 11) System should be supplied with 3 years warranty
- 12) System should be CE/ISO approved.

Chetana