

Ref: C-CAMP/L-080/2022-23 (C)

October 27, 2022

ENQUIRY

Dear Sirs,

Please let us have your lowest quotation for the following:

| Sl.No | Item/s description   | Qty.     |
|-------|--|----------|
| 1     | <p><b><u>Specification of -30° Freezer</u></b></p> <p>System should be vertical standing and should have working volume of 700 Lts ± 2%</p> <p>The system should be certified for research grade as per norms of IS &amp; CE or equivalent and suitable for continuous operation</p> <p>The system should have Microprocessor based PID controller it should have LED display with button control panel and user-friendly control system</p> <p>The system should have direct cooling with manual defrost.</p> <p>The freezer should have double door.</p> <p>The system should have the temperature control range of -20°C to -30°C</p> <p>The system should have the Maximum cooling performance of -30°C (Ambient Temperature: 35°C)</p> <p>The system should be certified for research grade as per norms of IS &amp; CE or equivalent and suitable for continuous operation</p> <p>The system should have minimum 4 adjustable shelves with 50kg capacity for storage of sample and easy access.</p> <p>The freezer should have audio- visual alarms for low/ high temperature, door open and power failure.</p> <p>The system should have remote alarm contacts NO/NC</p> <p>The system should be interior SS chamber and exterior with painted steel to avoid corrosion and starches</p> <p>The system should have an option to mute the audible alarms.</p> <p>The system should have automatic restart function in case of powers failure and after door opening and closing.</p> <p>The Noise level should not be more that 45db</p> <p>The system should be supplied with drain hose.</p> <p>The system should have ON/OFF control for fan motor for energy saving</p> <p>The system should have one Access port for external sensor which should be located at the back of the freezer. It should be supplied with stabilizer and all other accessories required for proper functioning of freezer.</p> <p>The system should have Caster moving wheels with stopper for easy movement with levelling feet.</p> <p>The system should have RS232 communication port</p> <p>The system should be supplied with suitable temperature chart recorder</p> <p>The system should be high power efficient, have high grade thermal puf insulation for freezer wall thickness 65 mm</p> <p>The system should have suitable LED display for temperature.</p> <p>The system should have minimum 3-year warranty</p> <p>The system should be CE/ISO certified.</p> <p>The system should be as per the Quality management system:ISO9001, Medical Device Management System:ISO13485 and Environmental Management System:ISO14001</p> <p>The system should have efficiency and low power consumption.</p> <p>The Company must quote the freezer with all the suitable documents to justify the specification of product.</p> <p>System should be suitable for power rating 230v, 50Hz single phase. System should have battery backup for controller, display and alarm.</p> <p>System should be highly energy efficient.</p> <p>Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required). Certificate of calibration and inspection from factory to be supplied with system.</p> <p>The system should be installed by the trained Engineer and users training should be given on Application, operation and maintenance</p> <p>The service team should be in Bangalore. Contact details and office address of Engineer should be submitted with this offer.</p> <p>The past performance/service support in NCBS/inStem/CCAMP should be satisfactory and the technical evaluation will be done accordingly.</p> <p>The system should have provision for easy disposal of water after manual defrosting through the drainage hose which should be attached to the main unit.</p> <p>The system should have insulated door with gasket for sample protection and uniformity.</p> <p>The system should have the provision for Remote monitoring.</p> | 1<br>No. |

Note:

1. The quotation shall be submitted in a sealed envelope duly superscribed with the enquiry number, and the due date for ..... Ref No:..... The bids should be addressed and to be couriered (sent by post/courier) to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and CCAMP/NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS **16/11/2022 BY 5.30PM.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto CCAMP Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: 90% of the payment shall be made only after completion of entire order at site, subject to acceptance by the group head through SAI and all other relevant documents at site. The balance 10% will be released subject to receipt of PBG for 3% order value and installation, All claims shall cease as per the Limitation Act.

In case of Imported, 90% payment shall be made through Sightdraft on presentation of complete and clear shipping documents and balance 10% of the amount shall be released after the receipt, installation, commissioning and acceptance of the equipment, against submission of 3% PBG.

10. The successful bidder shall have to execute a performance B/G in accordance with the guarantee/warranty for 3% value of the order. This B/G has to be executed on an appropriate value of stamp paper in terms of a bank guarantee drawn on any Nationalised Bank or first class Foreign Bank and shall remain valid till the completion of the Defect Liability period/warranty period, with 6 months claim period.

11. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.


13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

14. Since we are a research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No.51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices.

15. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

16. Liquidity Damages: If the equipment/ items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.

17. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services

  
Chetana S. R.