

Ref: C-CAMP/L-026/19-20 (T)

19-08-2019

## ENQUIRY

Dear Sirs,

Please let us have your lowest quotation for the following:

Sl.No.	Item/s description	Qty.
1.	<p><b><u>Specifications for Ultrapure (Type I) Water Purification System</u></b></p> <ul style="list-style-type: none"> <li>➤ System should be stand-alone Type-I water purification system, suitable for Biological research application.</li> <li>➤ The system must accept Type II water as feed.</li> <li>➤ Type I water must be produced from a mixed bed of ion exchange resin &amp; activated carbon cartridge to remove ionic and organic contaminants.</li> <li>➤ Type I water produced from the above step should pass through a safely housed, low-pressure mercury vapor UV lamp made of ultrapure quartz with dual wavelength (185/254nm) to ensure bacterial destruction and organic molecule oxidation.</li> <li>➤ Built-in Total Organic Carbon (TOC) indicator with 0.5mL quartz cell and UV lamp to accurately measure real-time TOC from 1-999ppb.</li> <li>➤ Must prevent deterioration of water quality during non-use. Must be compatible to place the production unit on the bench, under the bench, or mounted on the wall.</li> <li>➤ Point of delivery (POD) unit must be adjustable height and rotating arm-adjustable to any glassware, volumetric dispensing from 250 ml till 10Liter, and with display monitor that shows real-time resistivity, TOC, level of water in reservoir, volume dispensed and other alarms.</li> <li>➤ POD, at the collection end, must have a 0.22-micron filter and options to accommodate UF cartridge.</li> <li>➤ POD must also be able to accommodate cartridges to remove volatile organics, very specific trace organics, and trace metals.</li> <li>➤ The system should be CE/ISO Certified</li> <li>➤ The system should be supplied with all the accessories required to function</li> <li>➤ The system should have 3 Year Warranty. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).</li> <li>➤ The offer should include the spares like cartridges, membranes etc for 3 years as optional item.</li> <li>➤ The company should compulsory have the service team in Bangalore and also provide the Contact details of the Engineer</li> <li>➤ Warranty 3 Years</li> </ul>	2 Nos
2	<p><b><u>Laboratory Water Purification System Type II Grade</u></b></p> <ul style="list-style-type: none"> <li>➤ System should be stand-alone Type-II water purification system, suitable for Biological research application.</li> <li>➤ System should be suitable to work with the water quality of up to 2000<math>\mu</math>S/cm conductivity, Fouling Index (Silt Density Index) &lt; 7, Total Chlorine &lt; 3 ppm, PH: 4-10, produce Ultrapure Water with 4 to 10M<math>\Omega</math>.cm @25 °C resistivity</li> <li>➤ The system should have the production capacity of 40liter/hour</li> <li>➤ System should have Pre Filtration unit to counter the Iron content in water, particulate load Pre-filtration: 10 Micron filter, 5 Micron filter and 3 Micron filter.</li> <li>➤ System should have Pretreatment Cartridge with Activated Carbon and 0.5 Micron Filter.</li> <li>➤ System should have High Quality Thin film composite membrane RO which should be capable of achieving 99% rejection of Ions and all dissolved organics.</li> <li>➤ System should have Cartridges with high recovery loop to reduce the wastage of feed</li> </ul>	2 Nos

	<p>water to drain. Also it should have conductivity cell before and after RO to measure performance of RO membrane. RO waste water recovery should be up to 50%.</p> <p>➤System should have the Electro De-ionization module (EDI Module), with mixed Bed Ion Exchange Resin, so that Regeneration of the Resins happens on application of Electric current. EDI (Electro Deionization) module should not require softening pre-treatment.</p> <p>➤System should have Auto diagnostic facility with Error NO and Alarm Code and real time clock to log reports with date and time to ensure complete traceability.</p> <p>➤System should have Automatic Cleaning, Rising, and Flush mode.</p> <p>➤System should have suitable screen and user friendly GUI to check all parameters related to purification, storage and distribution.</p> <p>➤System should be able to check measuring value before and after each filtration stage in order to measure performance of filtration unit.</p> <p>➤System should give an alert in order to plan inventory.</p> <p>➤System should have storage capacity 100 liter with material of construction suitable to store Type II Water.Storage tank should be controlled by water purification unit. Tank should have valve for drain and dispense.</p> <p>➤Type II water should pass through feed water specific cartridge for removal of trace contaminants.</p> <p>➤System should have UV light treatment with wavelength of 185 and 254nm.</p> <p>➤System should have built in TOC indicator with the ability for self-calibration and check curve display.</p> <p>➤System should have compatibility to place either on the bench, under the bench or on the wall.</p> <p>➤The system should be CE/ISO Certified</p> <p>➤The system should be supplied with all the accessories required to function.</p> <p>➤System should have warranty of 3 years.</p> <p>➤Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).</p> <p>➤The offer should include the spares like cartridges, membranes etc for 3 years as optional item</p> <p>➤The company should compulsory have the service team in Bangalore and also provide the Contact details of the Engineer</p> <p>➤Warranty 3 Years</p>	
--	---	--

Note:

1. The bids shall be enclosed in an envelope and due date sealed duly marked Tender for \_\_\_\_ Ref No: C-CAMP/L-026/19-20 (T). The bids should be addressed and to be couriered (sent by post/courier) to 'THE HEAD-PURCHASE'.
2. The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section.
3. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 09-09-2019 BY 5.30 PM.
4. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.
5. The validity of your quotation should be for 60 days from the due date.
6. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.
7. Your quotation should indicate delivery period & warranty period.
8. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto C-CAMP Stores.
9. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.
10. Payment: within one month after delivery & acceptance/satisfactory installation.
11. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.
12. Since we are a public funded research institution, we are exempted from paying Customs duty (Except advolerum duty of 5% + 2%Cess and 1% Cus Sec & High Edu. CESS vide Notification No.51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all



scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

13. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.
14. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties
15. **"Liquidated damages:** If the \_\_equipment/ items\_\_\_\_\_ as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.
16. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services

Yours faithfully  
For and on behalf of  
Centre for Cellular and Molecular Platforms

*For Shabeen*  
GKVK, Bellary Road, Bangalore 560 065. India  
Phone +91-80-23666344. Fax +91-80-23636662  
[shabeentaj@ncbs.res.in](mailto:shabeentaj@ncbs.res.in)

