

Centre for Cellular and Molecular Platforms

Ref: C-CAMP/L-013/2022-23 (C)

May 04, 2022

ENQUIRY

Dear Sirs,

Please let us have your lowest quotation for the following:

l.N	Item/s description	Qty
•	Technical Specification for Bacterial incubator	01
	The system should have the capacity of 100 liters ± 5%	
	The system should be microprocessor controlled Incubator. It should have LCD Display with easy icons and user-friendly control system.	
	Should have a Temperature Control Range: 10°C to +60°C ±0.2°C	
ľ	The system Should have Temperature Uniformity of ±0.5 C at 35 °C	
ľ	The system Should have suitable Temperature Sensor	
t	The system should have the Air circulation by a fan to maintain temperature uniformity The system should have the Over-temperature protection Device	
	Should have 40mm Thickness of insulation with PUF Insulation Material The internal material of construction should be seamless SS 316	
ŀ	The system should have an option to mute the audible alarms	
1	The system should have the Corrosion resistance stainless steel interior and epoxy-	
ľ	powder coated exterior.	
	Should have minimum 3 adjustable shelves	
	Electronic and mechanical components should be enclosed, protected from accidental spills.	
	It should have automatic restart after power interruption with non-volatile memory	
ľ	The system should operate at 230V, 50 Hz AC power.	
	The system should have Noise Level less than 40 Db	
ľ	The system should be supplied with all the accessories required to function.	
	Compliance to each of the above points should be separately indicated and evidence	
	presence for each of them (Product brochures should be highlighted wherever required).	
ľ	The System should be CE/ISO Approved	
ŗ	The system should be power efficient and low power consumption.	
- 1	Compliance to each of the above points should be separately indicated and evidence	
	presence for each of them (Product brochures should be highlighted wherever required).	
	The system should be supplied with 5-year warranty.	
1	The service provider should have the base in Bangalore. Office address and service	
- 1	Engineer contact details should be enclosed with the quotation	
	The past performance/service support in NCBS/instem/ccamp should be satisfactory	
	and the technical evaluation will be done accordingly.	

Note:

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE"

with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and CCAMP/NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 24/05/2022 BY 5.30PM.

- 3.QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.
- 4. The validity of your quotation should be for 60 days from the due date.
- 5.All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.
- 6. Your quotation should indicate delivery period & warranty period.
- 7.Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto CCAMP Stores.
- 8.If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.
- 9. Payment: within one month after delivery & acceptance/satisfactory installation.
- 10.Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.
- 11. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.
- 12. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties
- 13.Since we are a research institution, we are exempted from paying Customs duty (Except advolerum duty of 5% + 2%Cess and 1% Cus Sec & High Edu. CESS vide Notification No.51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices
- 14.If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.
- 15. CCAMP is a public funded research institute and is entitled to concessional rate of GST @ 5% for certain items supplied for research purpose vide notification no. 47/2017 and 45/2017 dated 14th Nov, 2017. The offer should be submitted after fully considering the above notification.
- 16. Liquidity Damages: If the equipment/ items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.
- 17. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services

Chetana S. R.

GKVK, Bellary Road, Bangalore 560 065. India Phone +91-80-23666344.Fax+91-80-23636662 chetana@ncbs.res.in . www.ccamp.res.in