



Sl.No	NOTES
1.	DO NOT SCALE THE DRAWING. READ THE DIMENSION ONLY.
2.	ALL DIMENSION ARE IN MM
3.	ANY DISCREPANCY OR AMBIGUITY SHALL BE BROUGHT TO THE NOTICE OF ARCHITECT INCHARGE BEFORE QUOTING.
4.	THIS DRAWING IS TO BE READ WITH RELEVANT SPECIFICATION
5.	ALL SIZES SHOULD BE CHECKED AS PER SITE CONDITION.

1. IR Lamp - 150 Watts
2. Blower : 1/5 HP, 1420 RPM
3. Mini- Pleat HEPA - 400mm x 400mm x 75mm -1no.
4. Perforated Platform
5. UV Lamp - 15 Watts
6. Door Handle
7. Front Swing type PC Door.
8. Drawers at bottom - 3nos.
9. MOC : IS 304 Grade Stainless Steel.
10. Noise Level 35-40 dB "A"
11. Heavy duty Lockable castors
12. Storage Capacity : Approx 7 Garments
13. Airflow Pattern : Vertical Downflow.

REV. No	DESCRIPTION	DATE

TITLE: **STERILE GARMENTS STORAGE CABINET FOR MEDTECH CLEAN ROOM @ 3RD FLOOR C-CAMP**

ARCHITECT : Savitha K.S	DEALT : Basil	DATE : 09.04.2021	SCALE : NTS
----------------------------	------------------	----------------------	----------------